



National Accreditation Board for  
Testing and Calibration Laboratories

**CERTIFICATE OF ACCREDITATION**

**SCIENTIFIC AND INDUSTRIAL TESTING AND RESEARCH  
CENTRE**

has been assessed and accredited in accordance with the standard

**ISO/IEC 17025:2017**

**"General Requirements for the Competence of Testing &  
Calibration Laboratories"**

for its facilities at

83 & 84 AVARAMPALAYAM ROAD, K R PURAM POST, COIMBATORE, TAMIL NADU, INDIA

in the field of

**CALIBRATION**

Certificate Number: CC-4254

Issue Date: 31/01/2025

Valid Until: 30/01/2029

This certificate remains valid for the Scope of Accreditation as specified in the annexure subject to continued satisfactory compliance to the above standard & the relevant requirements of NABL.

(To see the scope of accreditation of this laboratory, you may also visit NABL website [www.nabl-india.org](http://www.nabl-india.org))

Name of Legal Entity: SCIENTIFIC AND INDUSTRIAL TESTING AND RESEARCH CENTRE

Signed for and on behalf of NABL



  
Anita Rani  
Director

  
N. Venkateswaran  
Chief Executive Officer



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

SCIENTIFIC AND INDUSTRIAL TESTING AND RESEARCH CENTRE, 83 & 84  
AVARAMPALAYAM ROAD, K R PURAM POST, COIMBATORE, TAMIL NADU, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-4254

**Page No**

57 of 57

**Validity**

31/01/2025 to 30/01/2029

**Last Amended on**

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)( $\pm$ )
138	THERMAL-TEMPERATURE	Temp Indicator with sensor of Freezer, Deep freezer, Refrigerator, Chamber,bath,Industrial Incubator (non medical pupose only), Salt Spray chamber,Autoclave(non medical pupose only) (single position)	Using PRT with Indicator by Comparison Method	(-) 40 °C to 300 °C	0.61 °C
139	THERMAL-TEMPERATURE	Temperature indicator with sensor of Furnace, Dry Block, Temperature indicator with sensor of Hot Air oven (Single position)	Using PRT with Indicator by Comparison Method	300 °C to 600 °C	0.71 °C
140	THERMAL-TEMPERATURE	Temperature Indicator with sensor of Industrial Furnace ,Dry Block(Single position)	Using S Type thermocouple with Indicator by comparison method	>600 °C to 1000 °C	1.84 °C
141	THERMAL-TEMPERATURE	Temperature Indicator with sensor of Industrial Furnace, Dry Block (Single position)	Using S Type thermocouple with Indicator by comparison method	>1000 °C to 1200 °C	2.44 °C

\* CMCs represent expanded uncertainties expressed at approximately the 95% level of confidence, using a coverage factor of  $k = 2$ .